

Table 3.18. Accuracy for the determination of atorvastatin

Concentration (mg/ml)	Average area	RSD%	Assay%
0.5	163782	0.44	101.5
1.0	326694	0.19	100.3
1.5	429061	0.52	98.7