

In reverse phase, the stationary phase used will be a nonpolar hydrocarbons, waxy liquids, or bonded hydrocarbons (such as C18, C8, etc.) while mobile phase is a polar solvents or mixtures such as methanol-water or acetonitrile-water, and etc. the most polar component is eluted first while less polar solvent has higher eluent strength ( Stoye D and Freitag W, 1996).

### **1.5.2 Validation of HPLC Instrumentation**

According to international conference of harmonization guideline (ICH), FDA and USP the operation of HPLC must be validated and maintained cleaned.

### **1.6 Definition of parameters used in the validation of the analytical HPLC method**

(USP, 2007; ICH, 1994).

#### **1.6.1 Precision**

The precision of an analytical procedure is the degree of agreement among individual test results when the procedure is applied repeatedly to multiple sampling of a homogeneous sample. The precision of an analytical procedure is expressed as the standard deviation.

#### **1.6.2 System Precision**

Used to ascertain injection repeatability and system suitability. It ascertains the effectiveness of the operating system as a single system.

#### **1.6.3 Method Precision**

Used to ascertain analysis repeatability by evaluating a number of samples containing known amounts of analyte.