

conditions. At each cycle, samples should be frozen for at least 12 hours before they are thawed. The accuracy for QC low and high after 3 cycles is within the accepted range which is 85-115%, table 45 and 46.

Sensitivity

The protein direct precipitation procedure was specified and sensitive for candesartan, where both blank and zero samples that examined from six different lots of plasma were attained the required clean chromatogram for specific method.