

Table 18: Intra-day precision and accuracy data for LLOQ samples of candesartan based on the standard calibration curve of day two validation.

Theo. Conc	AUC Drug	AUC IS	Ratios	Measured Conc.	Precision %	Accuracy%
10 QC Low	3589	54526	0.066	10.838	3.993	108.38
	3466	52636	0.066	10.846		108.46
	3304	51704	0.064	10.309		103.09
	3536	55152	0.064	10.366		103.66
	3430	51787	0.066	10.952		109.52
	3209	51478	0.062	9.870		98.70
Mean	3422.333	52880.500	0.065	10.530		105.302
STD	142.953	1579.498	0.002	0.421		4.205
CV%	4.177	2.987	2.343	3.993		3.993

Table 19: Intra-day precision and accuracy data for QC low samples of Candesartan based on the standard calibration curve of day two validation.

Theo. Conc.	AUC Drug	AUC IS	Ratios	Measured Conc.	Precision%	Accuracy%
30 ng/ml QC Low	6775	50882	0.133	29.560	5.677	98.53
	7556	53323	0.142	31.941		106.47
	6912	51639	0.134	29.757		99.19
	6927	52266	0.133	29.389		97.96
	7051	48792	0.145	32.722		109.07
	6776	52876	0.128	28.171		93.90
Mean	6999.500	51629.667	0.136	30.257		100.856
STD	291.757	1639.335	0.006	1.718		5.726
CV%	4.168	3.175	4.552	5.677		5.677