

Table 26: Inter day accuracy and precision for the quality control samples of Candesartan in the three days of validation.

Measured Concentration	10			30			500			800		
	Day One	Day Two	Day Three	Day One	Day Two	Day Three	Day One	Day Two	Day Three	Day One	Day Two	Day Three
	10.713	10.838	10.888	27.402	29.560	32.878	527.377	471.878	451.491	833.078	775.249	795.406
	10.647	10.846	10.076	27.342	31.941	29.949	493.988	464.492	466.471	794.983	743.945	840.571
	10.056	10.309	10.230	27.266	29.757	30.470	527.567	489.005	481.583	857.122	755.539	808.173
	10.113	10.366	10.336	30.769	29.389	28.339	532.078	475.061	477.998	814.572	763.083	794.577
	10.132	10.952	10.273	30.397	32.722	29.356	470.512	472.934	494.143	818.531	762.347	802.411
	9.701	9.870	9.368	29.809	28.171	30.865	466.683	466.292	499.045	841.007	734.964	771.060
Mean	10.317			29.799			484.922			794.812		
STD	0.437			1.696			23.591			35.549		
CV%	4.232			5.693			4.865			4.473		
Accuracy %	103.174			99.330			96.984			99.352		