

Table 14: Intra-day precision and accuracy data for LLOQ samples of candesartan based on the standard calibration curve of day one validation.

Theo. Conc.	AUC Drug	AUC IS	Ratios	Measured Conc.	Precision %	Accuracy %
10 QC Low	3717	57830	0.064	10.713	3.763	107.13
	3139	48985	0.064	10.647		106.47
	3702	59394	0.062	10.056		100.56
	3501	56022	0.062	10.113		101.13
	3051	48768	0.063	10.132		101.32
	3126	51014	0.061	9.701		97.01
Mean	3372.667	53668.833	0.063	10.227		102.270
STD	304.108	4661.220	0.001	0.385		3.848
CV%	9.017	8.685	1.814	3.763		3.763

Table 15: Intra-day precision and accuracy data for QC low samples of candesartan based on the standard calibration curve of day one validation.

Theo. Conc.	AUC Drug	AUC IS	Ratios	Measured Conc.	Precision %	Accuracy %
30ng/ml QC Low	6062	53322	0.114	27.402	5.778	91.34
	5807	51164	0.113	27.342		91.14
	6389	56403	0.113	27.266		90.89
	6385	51638	0.124	30.769		102.56
	6018	49110	0.123	30.397		101.32
	6474	53593	0.121	29.809		99.36
Mean	6189.167	52538.333	0.118	28.831		96.103
STD	264.928	2494.593	0.005	1.666		5.552
CV%	4.281	4.748	4.181	5.778		5.778