

### **3.2.1.2 Precision**

#### ***Validation for days 1-3***

The highest coefficient of variation in predicted concentration (CV%) was obtained by LLOQ of 2.21% and 2.13% for both day 1 and 2, respectively, while it was 2.81% for QC high. By contrast, the lowest variability of errors was shown by QC mid, QC high and LLOQ of 0.71%, 0.67% and 0.98% for day 1, 2 and 3, respectively (**Tables 3.25, 3.26 and 3.27**).