

List of tables

Table captions	Page No.
Table 2.1 Spiked serum or Krebs buffer of PRN serial spiking samples.	46
Table 2.2 Spiked serum or Krebs buffer of PRN quality control samples.	46
Table 2.3 Quality control concentrations used for method validation	48
Table 3.1 Summary of chromatographic conditions applied.	62
Table 3.2 Intra-day accuracy & precision data for all QC samples LLOQ, QC low, QC mid, and QC high of propranolol based on the standard calibration curves of the first day of validation (n=6). Corresponding calibration curve used in the calculation of measured concentration of intra-day one is shown in table 3.13 .	65
Table 3.3 Intra-day accuracy & precision data for all QC samples LLOQ, QC low, QC mid, and QC high of propranolol based on the standard calibration curves of the second day of validation (n=6). Corresponding calibration curve used in the calculation of measured concentration of intra-day two is shown in table 3.14 .	66
Table 3.4 Intra-day accuracy & precision data for all QC samples LLOQ, QC low, QC mid, and QC high of propranolol based on the standard calibration curves of the third day of validation (n=6). Corresponding calibration curve used in the calculation of measured concentration of intra-day three is shown in table 3.15 .	67
Table 3.5 Inter-day accuracy & precision data for the three days of QC samples validation (LLOQ, QC low, QC mid, and QC high) of propranolol based on the standard calibration curves of the three days of validation (n=6).	70
Table 3.6 Propranolol QC low samples (150 ng/ml) results for freeze-thaw stability test (n=3).	71
Table 3.7 Propranolol QC high samples (2500 ng/ml) results for freeze-thaw stability test (n=3).	71